

Supplement

Manual Title: 80PK-3A I. S. Supplement Issue: 1
Print Date: Sept. 1989 Issue Date: 3/10
Revision/Date: 4, 4/02 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

© 2010 Fluke Corporation. All rights reserved. Printed in U.S.A. **FLUKE**®

Change #1, 54379

Under **Specifications**, change the following:

From: **Junction Accuracy:** (With respect to ANSI MC96.1):
 ± 2.2 °C (3.96 °F) over the range of 0 °C to 260 °C
(32 °F to 500 °F)

To: **Accuracy:** (With respect to ANSI MC96.1): ± 4.0 °C
(7.2 °F) over the range of 0 °C to 260 °C (32 °F to 500 °F)